



The **All Flow™ Filter** is unsurpassed in Bacterial/Viral Efficiency in the lung function analysis application.

## TECHNOSTAT® Filter Media Data Sheet T150

Material	Weight	Color	Material Composition
Technostat	150 g/m2	White	Blended Synthetic Fiber
Scrim	15 g/m2	White	Spunbond Polypropylene
Total Media Weight	165 g/m2		
Available Forms	Single or Doubled Cover webs Roll goods and Cut parts		

### Filtration Performance

NaCl Penetration at 32 LPM	<1.00%	Tested in accordance to TSI8130 NaCl .1 micron particle size
NaCl Efficacy at 32 LPM	>99.00%	Tested in accordance to TSI8130 NaCl .1 micron particle size
Pressure drop at 32 LPM	<1.1 mm H2O	Tested in accordance to TSI8130 pressure drop standard
Air Permeability	>200 CFM	Tested in accordance to Spec ASTM, D737
BFE Efficiency	>99.9992%	Tested in accordance MIL-M-36954C, tested by Nelson Labs
VFE Efficiency	>99.9970%	Tested in accordance MIL-M-36954C, tested by Nelson Labs

Please contact Alliance Tech Medical for additional information, **1.800.848.8923**.

PO Box 6024 / Granbury, TX 76049  
1.800.848.8923 / 817.326.6357  
Fax: 817.326.2182

